

# UNCLASSIFIED

AD NUMBER
AD455002
NEW LIMITATION CHANGE
TO Approved for public release, distribution unlimited
FROM Distribution authorized to U.S. Gov't. agencies and their contractors; Administrative/Operational Use; 21 AUG 1964. Other requests shall be referred to Army Medical Research Lab., Fort Knox, KY.
AUTHORITY
USAMRL ltr, 26 Feb 1970

THIS PAGE IS UNCLASSIFIED

UNCLASSIFIED

AD\_4 5 5 0 0 2

DEFENSE DOCUMENTATION CENTER

FOR

SCIENTIFIC AND TECHNICAL INFORMATION

CAMERON STATION - ALEXANDRIA, VIRGINIA



UNCLASSIFIED

NOTICE: When government or other drawings, specifications or other data are used for any purpose other than in connection with a definitely related government procurement operation, the U. S. Government thereby incurs no responsibility, nor any obligation whatsoever; and the fact that the Government may have formulated, furnished, or in any way supplied the said drawings, specifications, or other data is not to be regarded by implication or otherwise as in any manner licensing the holder or any other person or corporation, or conveying any rights or permission to manufacture, use or sell any patented invention that may in any way be related thereto.

CATALOGED BY DDC  
AS AD NO. 455 602

# ARMY MEDICAL RESEARCH LABORATORY


FORT KNOX, KENTUCKY

REPORT NO. 610

EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND RAPID  
CHANGES IN PRESSURE UPON STAPEDECTOMIZED MONKEYS

Major John L. Fletcher, MSC  
Captain George D. Roberson, MC  
Michel Loeb, Ph. D.

45502



UNITED STATES ARMY  
MEDICAL RESEARCH AND DEVELOPMENT COMMAND

21 August 1964

Report Submitted 12 August 1964

Authors

Major John L. Fletcher, MSC (Ph. D.)	Department of Sensory Processes Division of Psychology USAMRL
Captain George D. Roberson, MC*	Chief, ENT Clinic Ireland Army Hospital
Michel Loeb, Ph. D.	Department of Sensory Processes Division of Psychology USAMRL

\*Now in practice at the Nalle Clinic, Charlotte, North Carolina.

The animals used in this study were handled in accordance with the "Principles of Laboratory Animal Care" established by the National Society for Medical Research.

Qualified requestors may obtain copies of this report from DDC.

Foreign announcement and dissemination of this report by DDC is limited.

REPORT NO. 610

EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND RAPID  
CHANGES IN PRESSURE UPON STAPEDECTOMIZED MONKEYS

Major John L. Fletcher, MSC  
Captain George D. Roberson, MC  
Michel Loeb, Ph.D.

Department of Sensory Processes  
Division of Psychology  
US ARMY MEDICAL RESEARCH LABORATORY  
Fort Knox, Kentucky

and

IRELAND ARMY HOSPITAL  
Fort Knox, Kentucky

21 August 1964

This Research Was Done Under

Traumatic Origins of Hearing Loss  
Subtask No. 02  
Military Psychophysiological Studies  
Task No. 01  
Military Psychophysiological Studies  
DA Project No. 3A012001A800

USAMRL Report No. 610  
DA Project No. 3A012001A800

ABSTRACT

EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND RAPID  
CHANGES IN PRESSURE UPON STAPEDECTOMIZED MONKEYS

OBJECT

To determine the effect of high intensity impulse noise and of rapid changes in pressure upon stapedectomized patients.

RESULTS

No experimental disarticulation of the prostheses was observed, nor were any behavioral manifestations of vestibular involvement seen. No significant difference between the two types of prostheses used was found.

CONCLUSIONS

There is no valid evidence in this study to support drastic duty limitation for stapedectomized personnel. No operational basis for choice between the polyethylene and vein graft procedure as opposed to the stainless steel piston could be found.

APPROVED: George S. Harker  
GEORGE S. HARKER, Ph. D.  
Director, Division of Psychology

APPROVED: Sven A. Bach  
SVEN A. BACH  
Colonel, MC  
Director

## EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND RAPID CHANGES IN PRESSURE UPON STAPEDECTOMIZED MONKEYS

Today we find the stapedectomy procedure the treatment of choice in the management of otosclerosis. The widespread incidence of otosclerosis, plus the popularity of the stapedectomy procedure, have resulted in a great number of persons undergoing this operation. The relative newness of the procedures and the difficulties attendant upon research in this area in part account for the dearth of research on the subsequent effects of this procedure upon the hearing abilities and susceptibilities of those so treated. Investigations have been made of the susceptibility of stapedectomized patients to noise induced temporary threshold shifts (1, 2, 3). So far, the findings appear to indicate no heightened susceptibility to acoustic insult. However, at best many problems beset the researcher in this endeavor. One critical problem is that of securing a proper control group against which to compare a stapedectomized population. Ideally, the control group should differ from the experimental group only in that they have not been stapedectomized. Certainly, auditory acuity of the two groups should be the same, because auditory acuity is a limiting factor in noise induced temporary threshold shifts (4). This requirement is exceedingly difficult to meet, and as a result, extreme caution must be observed in evaluating the results of studies where this condition was not met.

Several problems should be considered in evaluating the hazards of the environment to the stapedectomized person. Impulse noise, with its precipitous rise time and typically high overpressure, would logically be thought to pose more of a problem than would continuous noise, not only because of the possibility of noise induced hearing loss, but also because of the possibility of the inertia of the prosthesis-ossicular chain interacting with the steep rise time of the impulse to result in disarticulation of the prosthesis, or more drastic damage. Andersen *et al* (5) measured the transmission of sound before and after insertion of stapedial prostheses in cadaver temporal bones. They found no prosthesis weight-frequency related effects. They did, however, note that exposure to "violent" sounds resulted in disarticulation of the prostheses. They also observed that at such high levels (in excess of 100 dB) the prosthesis did not follow the movement of the incus. Subjective reports are frequently encountered suggesting that changes in altitude, such as those found in unpressurized or suddenly de-pressurized airplane cabins, result in failure of the prostheses. However, one might think that if the patient were able to equalize the pressure, i. e., "clear" his ears, danger from this source would be minimized.



in order to minimize activity during exposure (such as fighting, jumping, and other activities that might produce deleterious effects).

Group II, also 12 animals, 24 ears, was placed in an altitude chamber and ascended to a pressure equivalent altitude of 30,000 feet, then descended, free-fall, to 1,200 feet. This was repeated three times with 2 min intervals between runs. No sedation or tranquilization was given this group as we were afraid such treatment might interfere with the animals' ability to equalize pressure.

Group III, 11 animals and 22 ears, was housed with the other animals, treated just like them, but not exposed to gunfire or pressure change.

Immediately after exposure, the animals in Groups I and II were anesthetized and had their tympanic membranes reflected in order to determine the effects of the exposure upon the tympanic membrane, ossicular chain-prosthesis, middle ear mucosa, and the oval window reaction. At a later date the animals of the control group, Group III, were similarly scrutinized.

## RESULTS

The over-all results were surprisingly good from a clinical point of view, i.e., the prostheses were extremely resistant to disarticulation, so much so that none were experimentally interrupted. The observed reactions of the tympanic membrane, middle ear mucosa, and oval window were also remarkably mild. Both the PE + V and the SSP did well. The data do not seem to differentiate between the two as far as we can tell. Complete results of the experiment are presented in Tables 1, 2, 3, and 4. As can be seen, gunfire had no observable effect upon the prostheses and little observable effect upon the tympanic membrane, middle ear mucosa, and the oval window. Also, no behavioral effects of vestibular trauma were seen. Obviously, in view of the severity of the impulse noise to which these animals were exposed, more than a human would normally ever receive, we need not fear failure of the prostheses from this source. However, it is equally plain that we cannot, on the basis of this experiment, say anything about the damage to the hearing of the animal, only that the prostheses did not fail. In a further effort to promote failure of the prostheses by impulse noise, six animals of Group I were exposed to 12 rounds of gunfire from a 90 mm cannon. The average quasi-peak SPL, measured inside the cage, was in excess of 190 dB. Again, we failed to produce drastic results (see Table 4).

We could observe no significant basis for choice between the PE + V and SSP techniques. If basis for choice exists, it is surgical, not mechanical.

It is appropriate to mention here that more research is necessary before we can set exposure criteria for the patient's work environment. It will probably be some time before sufficient data are accumulated to enable us to evaluate the permanent effects of various noises upon the hearing of stapedectomized patients.

### CONCLUSIONS

1. High intensity impulse noise and rapid changes in air pressure do not appear sufficient to cause significant prosthesis failure in stapedectomized monkeys.
2. Of the two variables tested, within the limits of this experiment, pressure changes appear to be more traumatic.
3. No apparent difference appears to exist between the reliability of the stainless steel piston and the polyethylene and vein techniques.
4. Solely on a basis of fear of prosthesis failure, no reason can be seen for any drastic limitation of duty of stapedectomized patients.

### REFERENCES

1. Steffen, Ted N. and James C. Nixon. Stapedectomy and noise. *The Laryngoscope*, 73: 1044-1060, 1963.
2. Fletcher, John L. and William P. King. Susceptibility of stapedectomized patients to noise induced temporary threshold shifts. *Ann. Otol. Rhinol. Otolaryngol.* 72: 900-908, 1963.
3. McGee, T., M. Smith, and A. Graham. Stapedectomy patients in loud noise. Presented at Annual Meeting, American Academy of Ophthalmology and Otolaryngology, Las Vegas, 3 November 1962.
4. Glorig, A., W. D. Ward, and J. Nixon. Damage risk criteria and noise induced hearing loss. *Arch. Otolaryngol.* 67: 71-81, 1961.

## EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND RAPID CHANGES IN PRESSURE UPON STAPEDECTOMIZED MONKEYS

Today we find the stapedectomy procedure the treatment of choice in the management of otosclerosis. The widespread incidence of otosclerosis, plus the popularity of the stapedectomy procedure, have resulted in a great number of persons undergoing this operation. The relative newness of the procedures and the difficulties attendant upon research in this area in part account for the dearth of research on the subsequent effects of this procedure upon the hearing abilities and susceptibilities of those so treated. Investigations have been made of the susceptibility of stapedectomized patients to noise induced temporary threshold shifts (1, 2, 3). So far, the findings appear to indicate no heightened susceptibility to acoustic insult. However, at best many problems beset the researcher in this endeavor. One critical problem is that of securing a proper control group against which to compare a stapedectomized population. Ideally, the control group should differ from the experimental group only in that they have not been stapedectomized. Certainly, auditory acuity of the two groups should be the same, because auditor acuity is a limiting factor in noise induced temporary threshold shifts (4). This requirement is exceedingly difficult to meet, and as a result, extreme caution must be observed in evaluating the results of studies where this condition was not met.

Several problems should be considered in evaluating the hazards of the environment to the stapedectomized person. Impulse noise, with its precipitous rise time and typically high overpressure, would logically be thought to pose more of a problem than would continuous noise, not only because of the possibility of noise induced hearing loss, but also because of the possibility of the inertia of the prosthesis-ossicular chain interacting with the steep rise time of the impulse to result in disarticulation of the prosthesis, or more drastic damage. Andersen et al (5) measured the transmission of sound before and after insertion of stapedial prostheses in cadaver temporal bones. They found no prosthesis weight-frequency related effects. They did, however, note that exposure to "violent" sounds resulted in disarticulation of the prostheses. They also observed that at such high levels (in excess of 100 dB) the prosthesis did not follow the movement of the incus. Subjective reports are frequently encountered suggesting that changes in altitude, such as those found in unpressurized or suddenly de-pressurized airplane cabins, result in failure of the prostheses. However, one might think that if the patient were able to equalize the pressure, i. e., "clear" his ears, danger from this source would be minimized.

The Armed Forces are a particularly lucrative source of both problems and research possibilities with regard to stapedectomies. Gunfire noise exposure is, of course, typical of military personnel, and with the increasing use of aircraft of all kinds, and of airborne operations involving parachuting, a large number of persons are exposed to rapid changes in pressure. Problems of this nature suggested the research to be reported in the succeeding sections. Specifically, an investigation was made of the effects of impulse noise and sudden changes in pressure (simulating altitude changes) upon stapedectomized monkeys. Monkeys were used because of both the ethical and practical considerations involved.

#### METHOD

A total of 40 Cebus monkeys were procured and subjected to surgery. The Cebus monkey, a New World species, was used because its middle ear is quite accessible to surgery, unlike that of the Old World monkeys, and is markedly similar to that of man. Both ears of each animal were operated with the same type of prosthesis placed in each ear. Half the animals were treated with a modified Shea (6) technique utilizing a No. 90 polyethylene strut and vein graft (PE + V), while the other half had a stainless steel piston (SSP) made especially to fit the Cebus monkey. Five of the surgically completed animals expired subsequent to surgery from non-surgical causes so that our experiment was performed on a population of 35 animals. These animals were randomly divided into three groups. Group I was exposed to high intensity impulse noise on two occasions, to machine gun fire and to 90 mm cannon fire. Group II was exposed to rapid changes in pressure simulating changes in altitude, while Group III was the control group. They were operated, kept with and treated like the experimental groups except that they were not experimentally exposed to either of the variables.

Group I, which consisted of 12 animals (24 ears) was exposed to 2,000 rounds of fire from an M-73 7.62 cal machine gun. The fire was 5 sec bursts with a 5 sec interval between bursts. All animals exposed had had at least 2 months to recover from surgery. Firing was done on an open range in the field. The quasi-peak sound pressure level (SPL), measured inside the animal's retention cage, averaged 164 dB. Sound measurement was accomplished using a General Radio Model 1551-B sound level meter, a Massa 141-B microphone, and a General Radio Model 1556-B impact noise analyzer. Variability of the pressure was  $\pm 2$  dB. All animals were tranquilized 2 hours prior to exposure by injection of 1 mg/kg body weight of Seruylan (phencyclidine hydrochloride)

in order to minimize activity during exposure (such as fighting, jumping, and other activities that might produce deleterious effects).

Group II, also 12 animals, 24 ears, was placed in an altitude chamber and ascended to a pressure equivalent altitude of 30,000 feet, then descended, free-fall, to 1,200 feet. This was repeated three times with 2 min intervals between runs. No sedation or tranquilization was given this group as we were afraid such treatment might interfere with the animals' ability to equalize pressure.

Group III, 11 animals and 22 ears, was housed with the other animals, treated just like them, but not exposed to gunfire or pressure change.

Immediately after exposure, the animals in Groups I and II were anesthetized and had their tympanic membranes reflected in order to determine the effects of the exposure upon the tympanic membrane, ossicular chain-prosthesis, middle ear mucosa, and the oval window reaction. At a later date the animals of the control group, Group III, were similarly scrutinized.

## RESULTS

The over-all results were surprisingly good from a clinical point of view, i. e., the prostheses were extremely resistant to disarticulation, so much so that none were experimentally interrupted. The observed reactions of the tympanic membrane, middle ear mucosa, and oval window were also remarkably mild. Both the PE + V and the SSP did well. The data do not seem to differentiate between the two as far as we can tell. Complete results of the experiment are presented in Tables 1, 2, 3, and 4. As can be seen, gunfire had no observable effect upon the prostheses and little observable effect upon the tympanic membrane, middle ear mucosa, and the oval window. Also, no behavioral effects of vestibular trauma were seen. Obviously, in view of the severity of the impulse noise to which these animals were exposed, more than a human would normally ever receive, we need not fear failure of the prostheses from this source. However, it is equally plain that we cannot, on the basis of this experiment, say anything about the damage to the hearing of the animal, only that the prostheses did not fail. In a further effort to promote failure of the prostheses by impulse noise, six animals of Group I were exposed to 12 rounds of gunfire from a 90 mm cannon. The average quasi-peak SPL, measured inside the cage, was in excess of 190 dB. Again, we failed to produce drastic results (see Table 4).

The effects of altitude (pressure) as seen in Table 2, were somewhat more noticeable. For example, Table 1 shows that in Group I, the appearance of the tympanic membrane was normal for all subjects. In Group II, however, only eight of the membranes appeared normal; the rest varied from a slight to a generally injected appearance. Similarly, the middle ear mucosa was normal in Group I, while in Group II, only seven were normal, with the remaining 17 varying from a slight reaction to frank hemorrhage. The oval window reaction observed in the two groups was about the same. As we examine these data we are inescapably drawn to the conclusion that while neither of these experimental variables really wrecked havoc with the prostheses, the effects of the pressure change were considerably more marked. It is significant, we believe, that no failures of the prostheses were induced. The procedures are obviously highly reliable and quite resistant to forces considerably in excess of those one normally encounters in everyday life. This is not to say that failure will not occur, merely that we were unable to produce it. We desire to emphasize again, however, that these results related only to susceptibility of the prostheses to failure, and to tissue response, not to the effect of these variables upon the hearing of the stapedectomized animal.

We also believe that it is significant that no behavioral manifestations of vestibular damage could be observed. Any penetration of the vestibule by the prostheses should have produced immediate and observable behavioral changes. The fact that none were observed strongly suggests that no penetration occurred. Serial sections are being made of the temporal bones but results are not ready at this time.

Our results obviously do not support the findings of Andersen et al (5) who reported disarticulation of the prostheses at "violent" levels. This is not surprising, as their prosthesis was in an exposed dead temporal bone, while ours was in an intact living animal.

#### DISCUSSION

Based upon the results presented above, we can see no real need to drastically restrict the duty activities of stapedectomized persons after complete recovery from initial surgery. As always, common sense should be used but apparently the procedures are quite resistant to failure if successfully performed. However, as shown in Tables 1 and 2, original results of the insertion of the prostheses can be less than perfect, so that trauma could conceivably break down an initially imperfectly implanted prosthesis.

We could observe no significant basis for choice between the PE + V and SSP techniques. If basis for choice exists, it is surgical, not mechanical.

It is appropriate to mention here that more research is necessary before we can set exposure criteria for the patient's work environment. It will probably be some time before sufficient data are accumulated to enable us to evaluate the permanent effects of various noises upon the hearing of stapedectomized patients.

### CONCLUSIONS

1. High intensity impulse noise and rapid changes in air pressure do not appear sufficient to cause significant prosthesis failure in stapedectomized monkeys.
2. Of the two variables tested, within the limits of this experiment, pressure changes appear to be more traumatic.
3. No apparent difference appears to exist between the reliability of the stainless steel piston and the polyethylene and vein techniques.
4. Solely on a basis of fear of prosthesis failure, no reason can be seen for any drastic limitation of duty of stapedectomized patients.

### REFERENCES

1. Steffen, Ted N. and James C. Nixon. Stapedectomy and noise. *The Laryngoscope*, 73: 1044-1060, 1963.
2. Fletcher, John L. and William P. King. Susceptibility of stapedectomized patients to noise induced temporary threshold shifts. *Ann. Otol. Rhinol. Otolaryngol.* 72: 900-908, 1963.
3. McGee, T., M. Smith, and A. Graham. Stapedectomy patients in loud noise. Presented at Annual Meeting, American Academy of Ophthalmology and Otolaryngology, Las Vegas, 3 November 1962.
4. Glorig, A., W. D. Ward, and J. Nixon. Damage risk criteria and noise induced hearing loss. *Arch. Otolaryngol.* 67: 71-81, 1961.

5. Andersen, H. C., C. C. Hansen, and E. G. Neergaard. Experimental studies on sound transmission in the human ear. V. Function of stapedial prosthesis. *Acta Otolaryngol.* 57: 231-235, 1964.
6. Shea, J. J. Fenestration of the oval window. *Trans. Amer. Acad. Ophth. Otol.* 64: 611-627, 1960.



## LEGEND FOR TABLES

TM - Tympanic Membrane  
MM - Middle Ear Mucosa  
OW - Oval Window

### Tympanic Membrane

Normal (N) - Self-explanatory.  
Slight injury - Hyperemia in vascular strip area.  
General injury - Entire tympanic membrane hyperemic.

### Original Results

Self-explanatory.

### Middle Ear Mucosa

No reaction (NR) - Self-explanatory.  
Slight reaction - Generalized injection.  
Moderate reaction - Some petechiae.  
Frank hemorrhage - Self-explanatory.

### Oval Window

Slight - Very little reaction in the drum or the vein around the prosthesis.  
Moderate - Somewhat greater reaction around prosthesis.  
Marked - Drum plastered to incus with much reaction around prosthesis.

TABLE 1

## GROUP I - GUNFIRE

(Polyethylene Strut and Vein Graft)

Left Ear				Right Ear			
Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery	Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery
N	Originally slipped tube	NR	Slight	N	Good	NR	Slight
N	Good	NR	Moderate	N	Good	NR	Slight
N	Good	NR	Marked	N	Good	NR	Slight
N	Good	NR	Slight	N	Good	Moderate	Moderate
N	Good	NR	Slight	N	Good	NR	Slight
N	Good	NR	Slight	N	Good	NR	Slight
(Stainless Steel Piston)							
N	Good	NR	Slight	N	Good	NR	Slight
N	Good	NR	Slight	General injury	Good	Moderate	Moderate
N	Good	NR	Slight	N	Good	NR	Moderate
N	{Poor crimp, no slipping}	NR	Slight	N	Good	NR	Slight
N	Good	NR	Slight	N	Good	NR	Slight
N	Good	NR	Moderate	N	Good	NR	Slight

TABLE 2

## GROUP II - ALTITUDE

(No Perforations Seen in Any Ears)  
(Polyethylene Strut and Vein Graft)

Left Ear				Right Ear			
Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery	Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery
General injury	Good	Frank hem- orrhage	Slight	Slight	Good	Slight	Slight
N	Good	NR	Slight	Slight	Good	Slight	Slight
N	Good	NR	Slight	Slight	Good	Frank hem- orrhage	Slight
Slight	Good	Moderate	Slight	Slight	Good	NR	Slight
N	Good	NR	Slight	N	Slipped	NR	Slight
Slight	Slipped	Slight	Slight	Slight	Good	Frank hem- orrhage	Slight

(Stainless Steel Piston)

Slight	Good	Moderate	Slight	Slight	Good	Moderate	Moderate
Slight	Good	Moderate	Moderate	Slight	Good	NR	Moderate
Slight	Good	Frank hem- orrhage	Moderate	N	Poor crimp	NR	Slight
General	Good	Moderate	Moderate	Slight	Good	Moderate	Marked
injury	Good	Moderate	Slight	N	Good	Slight	Slight
Slight	Good	Slight	Slight	N	Good	NR	Slight
N							

TABLE 3  
GROUP III - CONTROLS\*  
(Polyethylene Strut and Vein Graft)

Left Ear				Right Ear			
Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery	Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery
N	OK	OK	OK	N	OK	OK	OK
N	OK	OK	OK	N	Slipped immedi- ately post-op	OK	OK
N	OK	Moderate	OK	N	OK	Moderate	OK
N	OK	OK	OK	N	Poor op- erative connec- tion	OK	OK
N	OK	OK	Moderate	N	OK	OK	OK
N	OK	OK	Moderate	N	OK	OK	OK
N	OK	OK	OK	N	OK	OK	Moderate
(Stainless Steel Piston)							
N	Good	NR	N	N	Riding high, poor crimp	NR	N
N	OK			N	OK	OK	OK
N	OK	OK	OK	N	OK	OK	OK
N	OK	OK	OK	N	OK	OK	OK

\*Some surgically completed animals expired from non-surgical causes, so control group slighted to allow correct experimental groups.

TABLE 4  
 EXPERIMENTAL - GUNFIRE 2 (90 mm)  
 (Polyethylene Strut and Vein Graft)

Left Ear				Right Ear			
Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery	Post-Exposure Appearance of the TM	Original Results	Reaction of the MM	Reaction of the OW, Due to Surgery
(Stainless Steel Piston)							
						Marked*	

\*With this one exception, gunfire produced no observable effect.

## ARMED FORCES

### DISTRIBUTION

Armed Forces Institute of Pathology, Washington, D. C.  
Army Attache, Fleet Post Office Box 79, Navy 100, New York, N. Y.  
Chief, Life Sciences Division, Office of the Commanding General, Washington, D. C.  
Chief, Medical Gen Lab, 40S, Japan, US Army Med Coad, APO 343, San Francisco, Calif.  
Chief, Preventive Med Div, Directorate of Professional Svc, SGO, Washington, D. C.  
Chief, Research and Development Command, Scientific Information Br, Washington, D. C.  
Cordt, Brooke Army Med Center, Med Field Svc School, Stinson Library, Ft. Sam Houston, Tex.  
CG, I Corps Group, APO 358, San Francisco, Calif.  
CG, Second US Army, Ft. George G. Meade, Md.  
CG, 7th Logistical Command, APO 612, San Francisco, Calif.  
CG, Eighth US Army, Attn: Surgeon, APO 301, San Francisco, Calif.  
CG, US Army Alaska, APO 949, Seattle, Wash.  
CG, US Army Chemical Research and Development Laboratories, Edgewood Arsenal, Md.  
CG, US Army Europe, APO 403, New York, N. Y.  
CG, US Army Hawaii, APO 957, San Francisco, Calif.  
CG, US Army Material Command, ATTN: AMC-RD, Washington, D. C.  
CG, US Army Material Research Lab, Biomechanical Lab, SGO, Washington, D. C.  
CG, US Army Medical Research and Development Command, Washington, D. C.  
CG, US Army Munitions Command, CBR Center, Edgewood Arsenal, Md.  
CG, US Army Natick Laboratories, (M1-7855) Natick, Mass.  
CG, US Army Ordnance Tank Auto Command, Detroit, Mich.  
CG, US Army Pacific Medical Office, APO 598, San Francisco, Calif.  
CG, US Army Quartermaster R&E Command, Technical Library, Natick, Mass.  
CG, US Army Ryukyu Islands, IX Corps, APO 331, San Francisco, Calif.  
CG, US Army Transportation R&E Command, Ft. Eustis, Va.  
CG, US Continental Army Command, Medical Section, Ft. Monroe, Va.  
CO, Brooke Army Medical Center, Surgical Research Unit, Ft. Sam Houston, Tex.  
CO, Chemical Corps, Biological Documents Laboratory, Ft. Detrick, Md.  
CO, Medical Technical Library, Fitzsimons General Hospital, Denver, Colo.  
CO, Picatinny Arsenal, Dover, N. J.  
CO, Quartermaster R&E FEA, US Army, Attn: Technical Publications Br, Ft. Lee, Va.  
CO, US Army Aeromedical Research Unit, US Army Aviation Center, Ft. Rucker, Ala.  
CO, US Army Combat Development Command, Combined Arms Group, Ft. Leavenworth, Kan.  
CO, US Army Combat Development Command, Materiel Svc, Ft. Sam Houston, Tex.  
CO, US Army Combat Development Command, Quartermaster Corps, Ft. Lee, Va.  
CO, US Army Combat Development Command, Engineering Command, Ft. Belvoir, Va.  
CO, US Army Command & General Staff College Library, Ft. Leavenworth, Kan.

## ARMED FORCES Cont.

### DISTRIBUTION

CO. US Army Electronic Proving Ground, Ft. Huachuca, Ark.  
CO. US Electronics R&D Laboratories, Attn: SELRA-GDA, Ft. Monmouth, N. J.  
CO. US Army Engineer R&D Command, STINFO Br. Ft. Belvoir, Va.  
CO. US Army Environmental Hygiene Agency, Library Br. Edgewood Arsenal, Md.  
CO. US Army Medical Liaison Br. Attn: Surgeon, Gorgas Hosp. Balboa Hts. Canal Zone  
CO. US Army Medical Res & Nutrition Lab, Fitzsimons General Hosp., Denver, Colo.  
CO. US Army Medical Research Unit, Ft. Clayton, Canal Zone  
CO. US Army Medical Research Unit, Inst for Medical Research Kuala Lumpur, Malaya  
CO. US Army Medical Research Unit, Ft. Detrick, Md.  
CO. US Army Medical Research Unit, APO 180, New York, N. Y.  
CO. US Army Missile Support Command, Attn: SMIDW-HO, Redstone Arsenal, Ala.  
CO. US Army Ordnance Arsenal, Frankford, Philadelphia, Pa.  
CO. US Army Personnel Research Office, Washington, D. C.  
CO. US Army Research Inst of Environmental Medicine, Natick Labs, Natick, Mass.  
CO. US Army Tropical Research, APO 851, New York, N. Y.  
CO. US Aviation Human Research Unit, US Continental Army, Ft. Rucker, Ala.  
Director of Research, US Army Air Defense, Human Research Unit, Ft. Bliss, Texas  
Director of Research, US Army Infantry, Human Research Unit Library, Ft. Benning, Ga.  
Director of Research, US Army Leadership, Human Res Unit Library, Presidio of Monterey, Calif.  
Director, US Army SEATO, Medical Research Laboratory, APO 146, San Francisco, Calif.  
Director, Walter Reed Army Inst of Res. Walter Reed Army Med Cent. Washington, D. C.  
Historical Univ, US Army Medical Service, Walter Reed Army Med Cent. Washington, D. C.  
Office of the Chief Psychi & Neuro Consultant, SGO, Washington, D. C.  
Office of the Chief R&D, Human Factors & Operations Research Div. Washington, D. C.  
Office of the Dir. Division of Neurophy, Walter Reed Army Inst of Res, Washington, D.C.  
Post Surgeon, US Army Dispensary, Springfield Armory, Springfield, Mass.  
Senior Medical Advisor, Hqs KNAG, APO 102, San Francisco, Calif.  
Walter Reed Army Inst of Res. Dept of Atomic Casualty Studies, Washington, D. C.

### HOSPITALS

CG, Brooke General Hospital Medical Library, Ft. Sam Houston, Tex.  
CG, Letterman General Hospital, Presidio of San Francisco, Calif.  
CG, Madigan General Hospital, Tacoma, Washington.  
CG, William Beaumont General Hospital Library, El Paso, Tex.  
CO, Bassett Army Hospital, Ft. Jonathan Wainwright, Fairbanks, Alaska.  
CO, DeWitt Army Hospital, Ft. Belvoir, Va.  
CO, Dunham Army Hospital, Carlisle Barracks, Pa.

ARMED FORCES - HOSPITALS Cont.

DISTRIBUTION

CO. Ireland Army Hospital, Ft. Knox, Ky.  
CO. Irwin Army Hospital, Ft. Riley, Kan.  
CO. Kenner Army Hospital, Ft. Lee, Va.  
CO. Kimbrough Army Hospital, Ft. George G. Meade, Md.  
CO. Kirk Army Hospital, Aberdeen Proving Ground, Md.  
CO. McDonald Army Hospital, Ft. Eustis, Va.  
CO. Martin Army Hospital, Ft. Benning, Ga.  
CO. Munson Army Hospital, Ft. Leavenworth, Kan.  
CO. Noble Army Hospital, Ft. McClellan, Ala.  
CO. Patterson Army Hospital, Ft. Monmouth, N. J.  
CO. Walson Army Hospital, Ft. Dix, N. J.  
CO. Walter Reed General Hospital, Washington, D. C.  
CO. Womack Army Hospital, Ft. Bragg, N. C.  
CO. US Army Hospital, Ft. Benjamin Harrison, Indianapolis, Ind.  
CO. US Army Hospital, Ft. Campbell, Ky.  
CO. US Army Hospital, Ft. Carson, Colorado.  
CO. US Army Hospital, Ft. Dix, N. J.  
CO. US Army Hospital, Ft. Devens, Mass.  
CO. US Army Hospital, Ft. Gordon, Ga.  
CO. US Army Hospital, Ft. Hood, Tex.  
CG. US Army Hospital, Ft. Jackson, S.C.  
CO. US Army Hospital, Ft. Jay, N. Y.  
CO. US Army Hospital, Ft. Leonard Wood, Mo.  
CO. US Army Hospital, Ft. McPherson, Ga.  
CO. US Army Hospital, Ft. MacArthur, Calif.  
CO. US Army Hospital, Ft. Monroe, Va.  
CO. US Army Hospital, Ft. Ord, Calif.  
CO. US Army Hospital, Ft. Polk, La.  
CO. US Army Hospital, Ft. Rucker, Ala.  
CO. US Army Hospital, Ft. Sill, Okla.  
CO. US Army Hospital, Ft. Stewart, Ga.  
CO. US Army Hospital, Ft. Wolters, Tex.  
CO. US Army Hospital, Dugway Proving Ground, Utah.  
CO. US Army Hospital, Redstone Arsenal, Ala.  
CO. US Army Hospital, Sierra Army Depot, Herlong, Calif.  
CO. US Army Hospital, White Sands Missil Range, N. Mex.



## ARMED FORCES - NAVY

### DISTRIBUTION

Aviation Medical Acceleration Lab, US Naval Air Development Center, Johnsville, Pa.  
Aviation Psychol Laboratory, US Naval School of Aviation Medicine, Pensacola, Fla.  
Bureau of Naval Weapons, Washington, D. C.  
Chief, Bureau of Yards and Docks, Washington, D. C.  
Chief of Naval Air Research Training, Glenview, Ill.  
Chief of Naval Air Technical Training, US Naval Air Station (75), Memphis, Tenn.  
Commander, Naval Missile Center, Technical Library, Point Mugu, Calif.  
CO, Naval Air Materiel Center, Technical Library, Philadelphia, Pa.  
CO, US Naval Civil Engineer Lab, Port Heunee, Calif.  
CO, US Naval Medical Field Research Lab, Library, Camp Lejeune, N. C.  
CO, US Naval Ordnance Test Station, Station Hospital, China Lake, Calif.  
CO, US Naval Radiological Defense Lab, San Francisco, Calif.  
CO, US Naval Weapons Station, Attn: QE Laboratory, Concord, California  
Director, Aerospace Crew Equipment Lab, Naval Air Engr Center, Philadelphia, Pa.  
Director Research Division (71) Bureau of Medicine & Surgery, Washington, D. C.  
Director, US Naval Research Laboratory, Code 2027, Washington, D. C.  
Director of Laboratories NMS, National Naval Medical Center, Bethesda, Md.  
Office of Naval Res Br, Document & Tech Infor, Box 39, Navy 100, Fleet PO, New York, N.Y.  
Office of Naval Research, Code 454, Washington, D. C.  
Officer in Charge, Exper Diving Unit, US Naval Sta, Navy Yard Annex, Washington, D. C.  
Officer in Charge, Operations Evaluations Group, Washington, D. C.  
Special Assistant, Medical Allied Sciences, Washington, D. C.  
Technical Ref. Library, Naval Med Res Inst, Nat'l Naval Med Center, Bethesda, Md.  
US Naval Research Laboratory, US Naval Submarine Base, New London, Groton, Conn.  
US Naval Supply Research & Development Facility, Library, Bayonne, New York  
US Naval Medical Neuropsychiatric Research Unit, San Diego, Calif.

### US AIR FORCE

Assistant for Ground Safety, US Air Force, Washington, D. C.  
Comdr, Rome Air Development Center, Griffiss AFB, N. Y.  
Comdr, 6570 AMRL (MRBA) Wright-Patterson AFB, Ohio  
Comdr, 6570 AMRL (MRBBH) Wright-Patterson AFB, Ohio  
Comdr, 6570 AMRL (MRBSL) Wright-Patterson AFB, Ohio  
Comdr, Wilford Hall, US Air Force Hospital, Lackland AFB, Texas  
Electronic Systems Div, Division Sci Lab, L. G. Hanscom Field, Bedford, Mass.  
Hqs-AF Office of Scientific Research, (Tempo Bldg D) Washington, D. C.

## ARMED FORCES - US AIR Cont.

### DISTRIBUTION

Langley Research Center, National Aeronautics & Space Admin, Hampton, Va.  
Office of the Surgeon, Air Tng Command, Randolph AFB, Tex.  
Office of the Surgeon, Hqs Strategic Air Command, Offutt, AFB, Neb.  
School of Aerospace Medicine, Aeromedical Library, Brooks AFB, Texas  
Science Division, AFRSTA, Hqs US Air Force, The Pentagon, Washington, D. C.  
US Air Force Epidemiological Laboratory, US Air Force, Lackland AFB, Tex.  
6571st Aeromedical Research Laboratory Library, Holloman AFB, N. Mex.

### GOVERNMENTAL AGENCIES

Argonne National Laboratory, Library Service Department, Argonne, Illinois  
Armed Forces Radiobiology Research Inst, National Naval Medical Center, Bethesda, Md.  
Brookhaven National Lab, Associated Universities Inc, Library, Upton, L. I., N. Y.  
Chief, Defense Atomic Support Agency, Document Library Branch, Washington, D. C.  
Chief, Radiation Br, National Cancer Institute, Bethesda, Md.  
Defense Documentation Center Hqs, Cameron Station, Alexandria, Va.  
Department of Health Educ & Welfare, Div of Radiol Health-PHS, Washington, D. C.  
Division of Biology & Medicine, US Atomic Energy Commission, Washington, D. C.  
Division of Medical Sciences, National Research Council, Washington, D. C.  
Executive Secretary, Committee of Vision, National Res Council, Washington, D. C.  
Federal Aviation Agency, CARI Library - AC-43.2, Oklahoma City, Oklahoma  
Library of Congress, Science & Technology Division, Washington, D. C.  
National Aeronautics & Space Administration, Washington, D. C.  
National Institute of Health, Division of Research Grants, Bethesda, Md.  
National Library of Medicine, Bethesda, Md.  
Office of Civil Defense, Technical Research Laboratory, Battle Creek, Mich.  
US Atomic Energy Commission, Reports Analysis Tech Infor Svc, Oak Ridge, Tenn.  
US Public Health Svc, Occupational Health Res & Tng, Facility, Cincinnati, Ohio

### OTHER AGENCIES

Abbott Laboratories, Science Information Services, N. Chicago, Illinois  
American Machine Foundry Co, New York, N. Y.  
Arctic Health Research Center Library, Anchorage, Alaska  
Central Medical Library, 11-42, The Boeing Co, Seattle, Washington  
Chief, Bioscience Section, Aerospace Division, The Boeing Co, Seattle, Washington  
Document Control Desk, The Biosearch Co, Boston, Mass.  
General Electric Co, Advanced Electric Center Library, Ithaca, N. Y.  
General Electric Co, Tempo Library, Santa Barbara, Calif.

#### OTHER AGENCIES Cont.

##### DISTRIBUTION

IBM Research Center, Engr Science Dept 475, Yorktown Heights, N. Y.  
ITT Federal Laboratories, Human Factors Group, Nutley, N. J.  
Kings County Hospital, Department of Anesthesiology Library, Brooklyn, N. Y.  
Lafayette Clinic 951 E. Lafayette Street, Detroit, Michigan  
Mayo Clinic, Director of Biophysics Division, Rochester, Minn.  
McDonnell Aircraft Corp, St. Louis, Mo.  
Psychological Abstracts, Washington, D. C.  
Rand Corporation Library, Santa Monica, California  
Rheem Califone Corporation, Los Angeles, Calif.  
Space Technical Laboratories, Subcommittee on Noise, Los Angeles, Calif.  
The Boeing Company Library, Military Aircraft Systems Division, Wichita, Kan.  
The John Crerar Library, Chicago, Illinois  
The Lovelace Foundation, Dept of Aerospace Med & Biobastronautics, Albuquerque, N. Mex.  
The Research Analysis Corp Library, McLean, Va.  
Yerkes Laboratories of Primate Biology, Orange Park, Fla.

##### MEDICAL COLLEGE/SCHOOL LIBRARIES AND DEPTS

Albany Medical College Library, Albany, N. Y.  
Bowman Gray School of Medicine, Library, Winston-Salem, N. C.  
Brown University, Library Documents Division, Providence, R. I.  
College of Medical Evangelists, Vernier Radcliff Memorial Library, Loma Linda, Calif.  
College of Physicians of Philadelphia, Library, Philadelphia, Pa.  
Columbia Univ, Dept of Psychology, New York, N. Y.  
Columbia Univ, Medical Library, New York, N. Y.  
Cornell Univ, Central Serial Record Department, Ithaca, N. Y.  
Cornell Univ, Medical Library, New York, N. Y.  
Creighton Univ, Medical Pharmacy Library, Omaha, Neb.  
Dartmouth College, Dana Biomedical Library, Hanover N. H.  
George Washington Univ, Office of Human Resources Research, Alexandria, Va.  
Harvard Univ, Medical Library, Boston, Mass.  
Indiana Univ, Medical Center Library, Indianapolis, Ind.  
Indiana Univ, School of Medicine Library, Bloomington, Ind.  
Jefferson Medical College, Library, Philadelphia, Pa.  
Johns Hopkins Univ, Welch Medical Library, Baltimore, Md.  
Kansas State Univ, Department of Psychology, Manhattan, Kan.  
Loma Linda Univ, White Memorial Medical Library, Los Angeles, Calif.

MEDICAL COLLEGE /SCHOOL LIBRARIES AND DEPTS Cont.

DISTRIBUTION

Marquette Univ. Medical-Dental Library, Milwaukee, Wis.  
Medical College of Virginia, Tompkins-McCaw Library, Richmond, Va.  
New York Academy of Medicine Library, New York, N. Y.  
New York Univ. College of Medicine, Research Division, New York, N. Y.  
New York Univ. College of Engr, Research Division, New York, N. Y.  
New York Univ. Medical Center Library, New York, N. Y.  
Northwestern Univ. Department of Psychology, Evanston, Ill.  
Northwestern Univ. Medical School Library, Chicago, Ill.  
Ohio State Univ. Chemical Abstracts Service, Columbus, Ohio  
Ohio State Univ. Engr Experiment Station, Systems Research Group, Columbus, Ohio.  
Ohio State Univ. Psycholinguistics Laboratory, Columbus, Ohio.  
Ohio State Univ. Research Center, Aviation Psychology Laboratory, Columbus, Ohio.  
Ohio State Univ. School of Optometry, Columbus, Ohio.  
Ohio State Univ. Tegar Library, Columbus, Ohio  
Purdue Univ. Department of Psychology, Lafayette, Ind.  
Rush Medical College Library, Chicago, Ill.  
St. Louis Univ. Medical School Library, St. Louis, Mo.  
Stanford Univ. Lane Medical Library, Palo Alto, Calif.  
State Univ of Iowa, College of Medicine Library, Iowa City, Iowa.  
State Univ of Iowa, Univ Hosp, Dept of Otolaryng & Maxillofacial Surg, Iowa City, Iowa.  
State Univ of New York, Anesthesiology Department, Brooklyn, N. Y.  
State Univ of New York, Medical Library, Downtown Medical Center, Brooklyn, N. Y.  
Texas Medical Center Library, Jesse H. Jones Library Bldg, Houston, Tex.  
Tufts Univ. Human Engineering Information Analysis Service, Medford, Mass.  
Tulane Univ. School of Medicine Library, New Orleans, La.  
Vanderbilt Univ. Hospital Library, Nashville, Tenn.  
Univ of Alabama, Medical Center Library, Birmingham, Ala.  
Univ of Arkansas, Medical Center Library, Little Rock, Ark.  
Univ of Buffalo, Department of Psychology, Buffalo, N. Y.  
Univ of Buffalo, Health Sciences Library, Buffalo, N. Y.  
Univ of California, Biomedical Library, Los Angeles, Calif.  
Univ of Chicago, Radiation Laboratory, Chicago, Ill.  
Univ of Cincinnati, College of Medicine, Kettering Lab, Cincinnati, Ohio  
Univ of Florida, College of Medicine, Gainesville, Fla.  
Univ of Georgia, Department of Psychology, Athens, Ga.

MEDICAL COLLEGE/SCHOOL LIBRARIES AND DEPTS Cont.

DISTRIBUTION

Univ of Illinois, Aeromedical Laboratory, Chicago, Ill.  
Univ of Illinois, Aviation Psychology Laboratory, Savoy, Ill.  
Univ of Illinois, Documents Division Library, Urbana, Ill.  
Univ of Illinois, Speech and Hearing Clinic, Champaign, Ill.  
Univ of Illinois, Training Research Laboratory, Urbana, Ill.  
Univ of Kansas, Glendening Medical Library, Kansas City, Kan.  
Univ of Kentucky, Psychology Department, Lexington, Ky.  
Univ of Louisville, Medical Library, Louisville, Ky.  
Univ of Maryland, Health Sciences Library, Baltimore, Md.  
Univ of Miami, School of Medicine, Department of Ophthalmology, Miami, Fla.  
Univ of Miami, School of Medicine, Department of Psychology, Coral Gables, Fla.  
Univ of Michigan, General Library of Serials & Documents, Ann Arbor, Mich.  
Univ of Minnesota, Serials Division Library, Minneapolis, Minn.  
Univ of Mississippi, Medical Center, Rowland Medical Library, Jackson, Miss.  
Univ of Missouri, Medical Library, Columbia, Mo.  
Univ of Nebraska, College of Medicine Library, Omaha, Neb.  
Univ of North Carolina, Div of Health Affairs, Memorial Hosp Library, Chapel Hill, Va.  
Univ of Oklahoma, Medical Center Library, Oklahoma City, Okla.  
Univ of Oregon, Medical School Library, Portland, Ore.  
Univ of Oregon, Department of Dermatology, Portland, Ore.  
Univ of Pittsburgh, Department of Psychology and Radiology, Pittsburgh, Pa.  
Univ of Pittsburgh, Falk Library, Pittsburgh, Pa.  
Univ of Rochester, Atomic Energy Project, Rochester, N. Y.  
Univ of Rochester, Department of Psychology, Rochester, N. Y.  
Univ of Rochester, School of Medicine & Dentistry, Rochester, N. Y.  
Univ of Rochester, Strong Memorial Hospital, Rochester, N. Y.  
Univ of Southern California, School of Medicine Library, Los Angeles, Calif.  
Univ of Southern California, Serials Library, Los Angeles, Calif.  
Univ of South Dakota, Medical Library, Vermillion, S. D.  
Univ of Tennessee, Mooney Memorial Library, Memphis, Tenn.  
Univ of Texas, The Radiobiology Laboratory, Austin, Tex.  
Univ of Texas, Medical Branch Library, Galveston, Tex.  
Univ of Texas, Speech & Hearing Clinic, Austin, Tex.  
Univ of Utah, Library of Medical Sciences, Salt Lake City, Utah.  
Univ of Vermont, College of Medicine Library, Burlington, Vt.  
Univ of Virginia, School of Medicine, Biostatistics, Charlottesville, Va.

MEDICAL COLLEGE/SCHOOL LIBRARIES AND DEPTS Cont.

DISTRIBUTION

Univ of Virginia, Medical Library, Univ Hospital, Charlottesville, Va.  
Univ of Virginia, Psychology Department, Charlottesville, Va.  
Univ of Washington, Health Sciences Library, Seattle, Wash.  
Univ of Wisconsin, Medical School Library, Madison, Wis.  
Univ of Wisconsin, Psychology Studies Division, Madison, Wis.  
Wayne State Univ, Medical Library, Detroit, Mich.  
West Virginia Univ, Medical Center Library, Morgantown, W. Va.  
Yale Univ, School of Medicine, Otologic Research Laboratory, New Haven, Conn.  
Xavier Univ, Department of Psychology, Cincinnati, Ohio

FOREIGN

Accessions Dept, Nat'l Lending Library for Sci & Tech, Boston SPA, Yorkshire, England  
American Embassy, The Foreign Svc of America, Attn: Asst Med Attache, London, England  
British Army Medical Liaison Officer, British Embassy, Washington, D. C.  
British Navy Staff Officer, Benjamin Franklin Sta, Attn: F. P. Ellis, Washington, D. C.  
Canadian Liaison Officer, Office of the Surgeon General, Washington, D. C.  
Chester Beatty Res Inst, Royal Cancer Hosp, Attn: Peter Alexander, London, England  
Commander P. H. Bonnel, Med En Chef De La, Marine, Paris, France  
Defence Research Member, Canadian Joint Staff, Washington, D. C.  
Dr. Hugo Arcadio, Hospital Militar Central, Asuncion, Paraguay.  
Dr. R. W. Brimblecomb, Chem Defence Expt, Est, Porton Down, Salisbury, Wilts, England.  
Dr. Martti J. Karvonen, Dir Physiol Dept, The Inst of Occup Health, Helsinki, Finland  
Dr. Jan Wersall, Dept of Histology, Karolinska Inst, Stockholm, Sweden  
Dr. G. H. Bradham, Dir Appl Physiol Lab, TUL & O.F.S. Chamber of Mines, Johannesburg, S. Africa  
Eichii Nagashima Extn Info Svc, The Nat'l Inst of Animal Health, Tokyo, Japan  
Escuela de Sanidad Militar, V. Edu Sanitar, Attn: Eugenio G. C. Ribeiro, Asuncion, Paraguay  
German Military Attache, Federal Ministry of Defense, Washington, D. C.  
Inst of Clinical Exper Surgery, Attn: Dr. Liska, Prague, Krc, Czechoslovakia  
Inst of Exper Med & Surg, Attn: Dr. Hans Selye, Dir Univ of Montreal, Montreal, Canada  
Inst of Physiology, Univ of Pisa, Attn: Guiseppe Moruzzi, Pisa, Italy  
Laboratory of General Pathology Therapy, Attn: Prof Zenon N. Bacq, Liege, Belgium  
Medical Sciences Library, University of Alberta, Edmonton, Alberta, Canada  
Oxford Univ, Department of Human Anatomy, Attn: Dr. A. R. Lind, Oxford, England  
Prof D. Bennati, Dir Inst de Ciencias Fisiologicas, Montevideo, Uruguay  
Prof R. W. Ditchburn, J. J. Thompson Physiol Lab, Berks Univ, Berks, England  
Prof D. M. MacKay, Univ of N. Staffordshir., Keele, Staffordshire, England  
Prof J. A. F. Stevenson Faculty of Med, Univ of Western Ont, London, Ont, Canada  
Royal Air Force Inst of Aviation Med, Attn: Dr. G. H. Byford, Farnborough, England  
Royal Air Force Inst of Aviation Med, Attn: Lt. J. C. Guignard, Farnborough, England  
Royal Society of Medicine Library, London England  
Univ of Western Ontario Med School, Attn: Dr. Allan C. Burton, London, Ont, Canada

AD Accession No.

US Army Medical Research Lab, Ft. Knox, Ky.  
EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND  
RAPID CHANGES IN PRESSURE UPON STAPEDECTO-  
MIZED MONKEYS - J. L. Fletcher, G. D.

UNCLASSIFIED

1. Audition
2. Stapedectomy
3. Noise Exposure

Report No. 610, 21 Aug 64, 11 pp & 1 - 4 tables - DA Project No.

3A012001A800, Unclassified Report  
In order to determine the effects of impulse noise and rapid changes in pressure upon stapedectomized patients, 40 Cebus monkeys were subjected to the stapedectomy procedure and later exposed to gunfire or rapid changes in pressure in an altitude chamber. Two different prostheses were used, half the monkeys receiving the polyethylene strut and vein graft, the other half getting a stainless steel piston prosthesis. Immediate post-exposure examination of the monkeys was made by reflecting the drums. No experimental disarticulation of the prostheses was observed, nor were there any behavioral manifestations of vestibular pathology. No significant differences were observed between the two different prostheses used. On the basis of this experiment, no valid reason for drastic duty limitation of stapedectomized patients can be seen.

AD

Accession No.

US Army Medical Research Lab, Ft. Knox, Ky.  
EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND  
RAPID CHANGES IN PRESSURE UPON STAPEDECTO-  
MIZED MONKEYS - J. L. Fletcher, G. D.

UNCLASSIFIED

1. Audition
2. Stapedectomy
3. Noise Exposure

Report No. 610, 21 Aug 64, 11 pp & 1 - 4 tables - DA Project No.

3A012001A800, Unclassified Report  
In order to determine the effects of impulse noise and rapid changes in pressure upon stapedectomized patients, 40 Cebus monkeys were subjected to the stapedectomy procedure and later exposed to gunfire or rapid changes in pressure in an altitude chamber. Two different prostheses were used, half the monkeys receiving the polyethylene strut and vein graft, the other half getting a stainless steel piston prosthesis. Immediate post-exposure examination of the monkeys was made by reflecting the drums. No experimental disarticulation of the prostheses was observed, nor were there any behavioral manifestations of vestibular pathology. No significant differences were observed between the two different prostheses used. On the basis of this experiment, no valid reason for drastic duty limitation of stapedectomized patients can be seen.

AD Accession No.

US Army Medical Research Lab, Ft. Knox, Ky.  
EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND  
RAPID CHANGES IN PRESSURE UPON STAPEDECTO-  
MIZED MONKEYS - J. L. Fletcher, G. D.

UNCLASSIFIED

1. Audition
2. Stapedectomy
3. Noise Exposure

Report No. 610, 21 Aug 64, 11 pp & 1 - 4 tables - DA Project No.

3A012001A800, Unclassified Report  
In order to determine the effects of impulse noise and rapid changes in pressure upon stapedectomized patients, 40 Cebus monkeys were subjected to the stapedectomy procedure and later exposed to gunfire or rapid changes in pressure in an altitude chamber. Two different prostheses were used, half the monkeys receiving the polyethylene strut and vein graft, the other half getting a stainless steel piston prosthesis. Immediate post-exposure examination of the monkeys was made by reflecting the drums. No experimental disarticulation of the prostheses was observed, nor were there any behavioral manifestations of vestibular pathology. No significant differences were observed between the two different prostheses used. On the basis of this experiment, no valid reason for drastic duty limitation of stapedectomized patients can be seen.

AD

Accession No.

US Army Medical Research Lab, Ft. Knox, Ky.  
EFFECTS OF HIGH INTENSITY IMPULSE NOISE AND  
RAPID CHANGES IN PRESSURE UPON STAPEDECTO-  
MIZED MONKEYS - J. L. Fletcher, G. D.

UNCLASSIFIED

1. Audition
2. Stapedectomy
3. Noise Exposure

Report No. 610, 21 Aug 64, 11 pp & 1 - 4 tables - DA Project No.

3A012001A800, Unclassified Report  
In order to determine the effects of impulse noise and rapid changes in pressure upon stapedectomized patients, 40 Cebus monkeys were subjected to the stapedectomy procedure and later exposed to gunfire or rapid changes in pressure in an altitude chamber. Two different prostheses were used, half the monkeys receiving the polyethylene strut and vein graft, the other half getting a stainless steel piston prosthesis. Immediate post-exposure examination of the monkeys was made by reflecting the drums. No experimental disarticulation of the prostheses was observed, nor were there any behavioral manifestations of vestibular pathology. No significant differences were observed between the two different prostheses used. On the basis of this experiment, no valid reason for drastic duty limitation of stapedectomized patients can be seen.

END

DTIC

6-86